

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I Yasuhiko Shinozaki, a Director of A&D Compnay,Limited,
 Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker^a A&D Compnay,Limited Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Manufacturer^b A&D Compnay,Limited Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Brand^c A&D Model^d UA-651SLPlus

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker^a A&D Compnay,Limited Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Manufacturer^b A&D Compnay,Limited Address 3-23-14 Higashi-ikebukuro Toshima-Ku,Tokyo 170-0013 JAPAN

Brand^c A&D Model^d UA-1020

Existing validated blood pressure measuring device.

which has previously passed the BHS protocol, the results of which were published as follows:

Wei-Fang Zeng, Yuan-Yuan Kang, Ming Liu, Yan Li, Ji-Guang Wang. Validation of the A&D UA-1020 upper-arm blood pressure monitor with six different-shaped or different-sized cuffs according to the British Hypertension Society Protocol.

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1–18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^e <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^g <input type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes:
- a Provide the name and address of the actual maker of the device.
 - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
 - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
 - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
 - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
 - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
 - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

5)The pressure sensor is replaced to a piezo electric sensor from an electrostatic capacitive sensor, but the accuracy of blood pressure measurement is equivalent between the two sensors.

6)The shapes of the connector are different

9)The equivalent device model name:UA-651SLPlus

10)Difference of case design. Both devices have the different casing.

11)The size and the symbols are different

13)Difference of Date and Time / Tricheck mode / Cuff Fit Error detection / Body Movement Error detection etc

14)UA-651SLPlus : 60 measurements, UA-1020 : 90 measurements

SECTION C Please check that the following are included with the application

A manual for the validated device

A manual for the device for which equivalence is being sought

Completed DET9 Form

An image of the device for which equivalence is being sought

An image of the screen layout of validated device*

An image of the screen layout of the device for which equivalence is being sought*

* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director Yasuhiko Shinozaki Company Stamp/Seal

Name Yasuhiko Shinozaki





Date 18 June 2019



Signature of Witness S. Ozaki

Name Shinobu Ozaki

Address 3-23-14 Higashi-ikebukuro Toshima-Ku, Tokyo 170-0013 JAPAN

Comparison of the AND UA-651SLPlus with the AND UA-1020

Devices – Item 9	AND UA-651SLPlus	AND UA-1020
Pictures		
Display Image		
Validation		BHS
Category	Upper Arm Blood pressure monitor	Upper Arm Blood pressure monitor
Casing – Item 10	<p><i>Dimensions</i> Approx : 96 【W】 × 68 【H】 × 130 【D】 mm</p> <p><i>Ports</i> Cuff port AC adaptor port</p> <p><i>Features</i> start Button Brand logo printing Model name printing SYS, DIA, PUL printing WHO Classification</p>	<p><i>Dimensions</i> Approx : 140 【W】 × 60 【H】 × 105 【D】 mm</p> <p><i>Ports</i> Cuff port AC adaptor port</p> <p><i>Features</i> start button / set button / ▲button / ▼button Brand logo printing Model name printing SYS, DIA, PUL printing WHO Classification</p>

Display – Item 11	<i>Type</i> liquid crystal display	<i>Type</i> liquid crystal display
Carrying/Mounting Facilities – Item 12	N/A	N/A
Software other than Algorithm – Item 13		Date and Time TriCheck™ mode MyPressure Setting mode (AUTO/180/210/240) Cuff Fit Error detection Body Movement Error detection
Memory Capacity Item 14	<i>Number of stored measurements</i> Last 60 measurements	<i>Number of stored measurements</i> Last 90 measurements
Printing Facilities Item 15	N/A	N/A
Communication Facilities – Item 16	N/A	N/A
Power Supply Item 17	4×1.5V batteries(R6P, LR6 or AA) or AC adapter(TB-233C) (optional)	4×1.5V batteries(R6P, LR6 or AA) or AC adapter(TB-233C) (optional)
Other differences	<p><i>Other Details on Equivalent device that are different to Validated device</i></p> <p><i>Sensors</i> Semiconductor sensor</p> <p><i>Cuff plug</i> Outline</p> 	<p><i>Other Details on Validated device that are different to Equivalent device</i></p> <p><i>Sensors</i> Capacitance sensor</p> <p><i>Cuff plug</i> Outline</p> 
Same Criteria	<p>Measurement</p> <p><i>Accuracy</i> Pressure: ±3 mmHg Pulse: ±5 %</p>	<p>Measurement</p> <p><i>Accuracy</i> Pressure: ±3 mmHg Pulse: ±5 %</p>

	<p><i>Method</i> Oscillometric measurement</p> <p><i>Ranges</i> Pressure: 0 - 299 mmHg Systolic pressure: 60 - 279 mmHg Diastolic pressure: 40 - 200 mmHg Pulse: 40 - 180 beats/minute</p> <p><i>Inflation</i> Automatic inflation</p> <p><i>Deflation</i> Automatic deflation</p> <p><i>Cuffs (Please state sizes and materials used)</i> 16-24cm, 22-32cm, 23-37cm, 31-45cm Nylon</p> <p><i>Sensors</i> Semiconductor sensor</p> <p><i>Measurement Records</i> SYS,DIA,PUL</p> <p><i>Measurements other than Blood Pressure</i> Heart rate WHO classification</p> <p>Buttons/Switches <i>Power</i> Start button</p> <p><i>Measurement Records</i> Memory recall button – Start button for 3sec</p> <p><i>Function</i> N/A</p> <p><i>Analysis</i></p>	<p><i>Method</i> Oscillometric measurement</p> <p><i>Ranges</i> Pressure: 0 - 299 mmHg Systolic pressure: 60 - 279 mmHg Diastolic pressure: 40 - 200 mmHg Pulse: 40 - 180 beats/minute</p> <p><i>Inflation</i> Automatic inflation</p> <p><i>Deflation</i> Automatic deflation</p> <p><i>Cuffs(Please state sizes and materials used)</i> 16-24cm, 22-32cm, 23-37cm, 31-45cm Nylon</p> <p><i>Sensors</i> Capacitance sensor</p> <p><i>Measurement Records</i> SYS,DIA,PUL</p> <p><i>Measurements other than Blood Pressure</i> Heart rate WHO classification</p> <p>Buttons/Switches <i>Power</i> Start button</p> <p><i>Measurement Records</i> Memory recall button – ▲button or ▼button</p> <p><i>Function</i> Date and time setting - set button</p> <p><i>Analysis</i></p>
--	---	--

	<p>N/A</p> <p><i>Event Marking</i> N/A</p> <p><i>Communication</i> N/A</p> <p><i>Display/Symbols/Indicators</i> <i>Preparation</i> Zero is blinking</p> <p><i>Measurement Procedure</i> Pressure value Heart mark Pressure bar indicator</p> <p><i>Post Measurement</i> Systolic blood pressure Diastolic blood pressure Pulse rate WHO classification</p> <p><i>Measurement Records</i> Systolic blood pressure Diastolic blood pressure Pulse rate WHO classification Memory mark symbol Memory number</p> <p><i>Date and Time</i> N/A</p> <p><i>Power</i> Battery detection symbol</p>	<p>N/A</p> <p><i>Event Marking</i> N/A</p> <p><i>Communication</i> N/A</p> <p><i>Display/Symbols/Indicators</i> <i>Preparation</i> Zero is blinking</p> <p><i>Measurement Procedure</i> Pressure value Heart mark Pressure bar indicator</p> <p><i>Post Measurement</i> Systolic blood pressure Diastolic blood pressure Pulse rate WHO classification Date and Time</p> <p><i>Measurement Records</i> Systolic blood pressure Diastolic blood pressure Pulse rate WHO classification Date and Time Memory mark symbol Memory number</p> <p><i>Date and Time</i> Year, Month, Day, Hour, Minute</p> <p><i>Power</i> Battery detection symbol</p>
--	--	---

	<p><i>Function</i> Average data</p> <p><i>Communication</i> N/A</p> <p><i>Features</i> N/A</p> <p><i>Not described</i> N/A</p> <p>Algorithms <i>Averages and Differences</i> N/A</p> <p><i>Diagnostic</i> N/A</p> <p><i>Functions</i> N/A</p> <p><i>Communication</i> N/A</p>	<p><i>Function</i> Average data TriCheck™ mode MyPressure Setting mode (AUTO/180/210/240) Cuff Fit Error detection Body Movement Error detection</p> <p><i>Communication</i> N/A</p> <p><i>Features</i> N/A</p> <p><i>Not described</i> N/A</p> <p>Algorithms <i>Averages and Differences</i> N/A</p> <p><i>Diagnostic</i> N/A</p> <p><i>Functions</i> N/A</p> <p><i>Communication</i> N/A</p>
Comparable Criteria		

Comments	Satisfactory explanations received for cuff connectors and sensor/transducer.
Recommendation	Recommended
Date	05 July 2019