

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2006

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items online.

I Takefumi Nakanishi Director of Omron Healthcare Europe B.V.
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Omron HEM-7101 (HEM-7101-SH)
Blood pressure measuring device for which validation is claimed

blood pressure measuring device and the

Omron M2 Compact (HEM-7102-E)
Existing validated blood pressure measuring device

blood pressure measuring device, which has previously passed the International protocol, the results of which were published as follows

Asmar R, Khabouth J, Topouchian J, El Feghali R, Mattar J

Authors(s)

Validation of three automatic devices for self-measurement of blood pressure according to the International Protocol: The Omron M3 Intellisense (HEM-7051-E), the Omron M2

Compact (HEM 7102-E), and the Omron R3-I Plus (HEM 6022-E)

Title

Blood Pressure Monitoring

2010; 15:49-54

Publication

Year Volume Pages

The only differences between the devices involve the following components:

(When a component is not relevant, both Yes and No should be left blank. Please provide details on any differences below.)

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	6	Cuff or Bladder	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	14	Memory Capacity/Number of stored measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>

Brief explanation of differences and further relevant details:

10) No AC adapter port.

11) No symbol for irregular heart beat. The symbol for mmHg/kPa is added.

13) No function to detect irregular heart beat. The function of switching mmHg/kPa is added.

14) Stores 21 readings instead of 14.

17) AC adapter is not available.



SECTION B - Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original along with manuals for both devices to our address below.

Signature of Director T. Nakanishi

Name Takefumi Nakanishi

Date 04 February 2010

Signature of Witness J. Meijer





Name Janet Meijer

Address Omron Healthcare Europe B.V., Kruisweg 577, 2132NA Hoofddorp, The Netherlands

Company Stamp/Seal

OMRON HEALTHCARE EUROPE B.V.
Kruisweg 577
NL-2132 NA Hoofddorp
P.O. Box 2150 NL- 2130 GL Hoofddorp
Tel. +31 - 20 354 82 00
Fax +31 - 20 354 82 01

Comparison of the Omron HEM-7101 (HEM-7101-SH) with the Omron M2_Compact (HEM-7102-E)

Devices	HEM-7101 (HEM-7101-SH)	M2_Compact (HEM-7102-E)
Pictures		
Display		
Validation		ESH
Device 1 Criteria	<p>Display/Symbols/Indicators Settings Current unit (kPa / mmHg) marker 11</p> <p>Algorithms Parameter Settings Unit conversion (kPa / mmHg) 13</p>	
Same Criteria	<p>Measurement Accuracy Pulse accuracy ± 5% 1, 5</p> <p>Method Oscillometric measurement method 1, 5 Pulse 40 bpm -180 bpm 1, 5, 8 Manually initiated measurements 13 Measurements are from single inflations 13</p> <p>Inflation Inflation 0 mmHg - 299 mmHg 1, 5, 7</p>	<p>Measurement Accuracy Pulse accuracy ± 5% 1, 5</p> <p>Method Oscillometric measurement method 1, 5 Pulse 40 bpm -180 bpm 1, 5, 8 Manually initiated measurements 13 Measurements are from single inflations 13</p> <p>Inflation Inflation 0 mmHg - 299 mmHg 1, 5, 7</p>

Automatic Inflation	7	Automatic Inflation	7
Manually adjustable inflation pressure	7	Manually adjustable inflation pressure	7
<i>Deflation</i>		<i>Deflation</i>	
Automatic Deflation	8	Automatic Deflation	8
Automatic safety release valve	8	Automatic safety release valve ^{Query 1}	8
<i>Cuffs</i>		<i>Cuffs</i>	
Medium 146 mm × 446 mm (Arm circ. 22 to 32 cm) ^{Query 2}	6	Medium 146 mm × 446 mm (Arm circ. 22 to 32 cm) ^{Query 2}	6
<i>Sensors</i>		<i>Sensors</i>	
Pressure sensor: capacitive	5	Pressure sensor: capacitive	5
Buttons/Switches		Buttons/Switches	
<i>Power</i>		<i>Power</i>	
On/Off with Start/Stop	10	On/Off with Start/Stop (O/I Label)	10
<i>Measurement Records</i>		<i>Measurement Records</i>	
Memory	10	Memory	10
Display/Symbols/Indicators		Display/Symbols/Indicators	
<i>Measurement Procedure</i>		<i>Measurement Procedure</i>	
Deflation symbol	11	Deflation symbol	11
During Measurement: BP Level & Heartbeat	11	During Measurement: BP Level & Heartbeat	11
<i>Post Measurement</i>		<i>Post Measurement</i>	
SBP, DBP and Pulse	11	SBP, DBP and Pulse	11
Measurement error EE , E , E/E and $E_{\alpha 25}$	11	Measurement error EE , E , E/E and $E_{\alpha 25}$	11
Hypertension (Blinking heartbeat)	11, 13	Hypertension (Blinking heartbeat)	11, 13
<i>Measurement Records</i>		<i>Measurement Records</i>	
Memory icon	11	Memory icon	11
<i>Power</i>		<i>Power</i>	
Low battery	11, 17	Low battery	11, 17
Algorithms		Algorithms	
<i>Diagnostic</i>		<i>Diagnostic</i>	
Normotension/Hypertension	13	Normotension/Hypertension	13
135 / 85 mmHg thresholds	13	135 / 85 mmHg thresholds	13
Case		Case	
<i>Display</i>		<i>Display</i>	
Single screen display	10	Single screen display	10
Segment LCD	10	Segment LCD	10
<i>Power</i>		<i>Power</i>	
4 “AAA” batteries ~ 300 measurements	17	4 “AAA” batteries ~ 300 measurements	17
Automatic switch-off when not used for 5 min	17	Automatic switch-off when not used for 5 min	17

Comparable Criteria	Measurement <i>Accuracy</i> BP accuracy ± 4 mmHg 1, 5 <i>Inflation</i> Press button if BP > 220 mmHg 7 <i>Measurement Records</i> Memory: 21 measurements 14	Measurement <i>Accuracy</i> BP accuracy ± 3 mmHg 1, 5 <i>Inflation</i> Press button if BP > 170 mmHg 7 <i>Measurement Records</i> Memory: 14 measurements 14
Device 2 Criteria		Measurement <i>Cuffs</i> Large (Arm circ. 32-42 cm) (Optional) 6 Display/Symbols/Indicators <i>Post Measurement</i> Irregular heartbeat 11, 13, 18 Algorithms <i>Diagnostic</i> Irregular heartbeat detection 13 Case <i>Power</i> AC adapter (Optional) 17
Web link		http://www.

Comments	<p>Query 1 Rapid pressure release: The manual, for the HEM-7101, include two deflation entries. In addition to the regular deflation, there is an automatic exhaust valve for rapid pressure release. This is understood to be a safety feature. It appears not to be available for the M2 Compact. There is no reference to this difference in the declaration. Please explain.</p> <p>Response 1 <i>The fact we have is that the M2 Compact (HEM-7102-E) and the HEM-7101 (HEM-7101-SH) have same deflation mechanism. They have same valves for deflation system, as you mentioned, which are the regular deflation valve (slow deflation during measurement) and the rapid exhaust valve (release pressure rapidly from air system in the device after measurement to make comfortable and safe patients). Also these 2 valves are operated by automatic. In some device's manual e.g. M3 Intellisense (HEM-7051-E), we mention only "Deflation: Automatic pressure release valve" as one function of automatic deflation so that we could provide easy explanation to end users.</i></p> <p>Query 2 There appear to be some differences in the cuffs supplied with the monitors.</p> <p>There are different part numbers between those listed for the devices. No part numbers is provided for the HEM-7101 and no difference is made in the declaration. It is taken that there are no changes.</p> <p>Response 2 These cuffs have no differences except cloth covers. The parts number difference comes from different cloth covers.</p>
-----------------	---

Recommendation	The queries were adequately answered. Further queries need to be made regarding the accuracy ranges and the BP above which a manually initiated boost is required. Equivalence is recommended subject to an adequate responses.
Date	26/08/2010